

## [1] TYPE EXAMINATION CERTIFICATE

- [2] Equipment intended for use in potentially explosive atmospheres - Directive 2014/34/EU
- [3] Type Examination Certificate number: **DTI 21ATEX0160X**
- [4]Equipment:Refrigerators and freezersTypeBioCompact II RR/RF 210, 310, 410 (see type key under item 15)
- [5] Manufacturer: Gram Commercial
- [6] Address: Aage Grams Vej 1 DK-6500 Vojens, Denmark
- [7] This equipment and acceptable variation thereof are specified in the schedule to this certificate and in the documents that this certificate refers to.
- [8] Danish Technological Institute certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to design and construction of equipment intended for use in potentially explosive atmospheres given in Annex II of the ATEX Directive.

The examinations and test results are recorded in confidential report no. 21ATEX0160-00, order no: 977822

- The essential Health and Safety Requirements are assured by compliance with: EN IEC 60079-0:2018/AC:2020, EN 60079-11:2012, EN 60079-15:2019, EN 60079-7:2015+A1:2018
- [10] The sign "X" placed after the certificate number, indicates that the equipment or protective system is subject to special conditions for safety use specified in the schedule to this certificate.
- [11] This certification relates to the design, examination and tests of the specified equipment or protective system. Further requirements of the directive may apply to the manufacturing process and supply of this equipment or protective system. These are not covered by this certificate.
- [12] The marking of the equipment shall include the following:

II 3G Ex ec nC ic IIB Tx Gc 2022-01-10 **Danish Technological Institute Certifications & Inspection** Jakob Nittegaard ATEX Manager

Kongsvang Allé 29 DK-8000 Århus C Denmark

Tel. +45 72 20 10 00 Fax +45 72 20 10 19 www.teknologisk.dk



# [13]Schedule[14]TYPE EXAMINATION CERTIFICATE

### DTI 21ATEX0160X

#### [15] **Description of equipment**

Gram BioLine BioCompact II range of refrigerators (RR) and freezers (RF) are designed and manufactured to provide safe and precise conditions for the items stored. The cabinets are designed for the following operating ranges RR +2  $^{\circ}C/+20 ^{\circ}C$ , RF -25  $^{\circ}C/-5 ^{\circ}C$ .

#### Variants covered by this certificate:

Cabinet	210								
Name	BioCompact II RR210 (10)			BioCompact II RR210 (11)			BioCompact II RF210 (10)		
	НС	н	FC	НС	HC HFC		HC	HFC	
Connection	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz
Temperature									
class	Т6	T5	T5	T6	T5	T5	Т3	Т3	T3

Cabinet	310								
Name	BioCompact II RR310 (10)			BioCompact II RR310 (11)			BioCompact II RF310 (10)		
	нс	н	FC	НС	HFC		HC	HFC	
Application	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz
Temperature									
class	T6	T5	T5	T6	T5	T5	Т3	Т3	Т3

Cabinet					410					
Cabinet	_	410								
Name	BioCom	BioCompact II RR410 (10)			BioCompact II RR410 (11)			BioCompact II RF410 (10)		
	нс	н	FC	нс	н	FC	НС	н	FC	
Application	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz	230V/50Hz	230V/50Hz	230V/60Hz	
Temperature										
class	Т6	T5	T5	Т6	T4	T5	Т3	Т3	Т3	

Specified ambient temperature (10 - solid door):  $+10 \degree C < t_{amb} < +35 \degree C$ Specified ambient temperature (11 - glass door):  $+10 \degree C < t_{amb} < +32 \degree C$ 

> This certificate must be reproduced in its entirety only and without any changes, including appendices. Project No.: 21.0160 Version: 00



Certification & Inspection

In addition to the BioCompact II models, the below listed accessories are covered by this certificate.

The accessories are specified by the type number system below.

BioCom	pact II
Temper	ature range (xx)
RR	Regular refrigerator +2/+20 °C
RF	Regular freezer -25/-5 °C
Refriger	5
Н	
G	R134a (only outside EU) R600a
Exterior	
C	Stainless
L	White Lacquered
Equipmo	
10	Solid door
-	
11 12	Glass door Customized door
12	RUKO lock (DK)
20	Right hinged self-closing door
20	
30	Left hinged self-closing door Wheels/castors H=125 mm
31	Castors, 360° castors, 2 w/brakes
33	Wheels/castors H=109 mm
35	Legs H=100/135 mm
35	Legs H=135/200 mm
38	Prepared for mounting on a plinth
39	Skids (only 210/310 units)
41	E-sensor
42	Customized sensor
81	Low temperature protection
82	60 Hz compressor
90	Combination kit
91	Wall mounting
V	# of liquid containers incl. holder
A	# of aluminium drawers
AA	# of aluminium drawers 300 mm
E	# of ABS drawer
M	# of wire baskets in white plastic coating
MM	# of wire baskets in white plastic coating 300 mm
P	# of perforated shelves
PP	# of perforated shelf in stainless steel 300 mm
T	# of wire shelves in stainless steel
TT	# of wire shelf in stainless steel 300 mm
W	# of wire shelves in white plastic coating
WW	# of wire shelves in white plastic coating 300 mm
Z	# of perforated shelves in white
ZZ	# of perforated shelves in white

#### [16] **Report number**

DTI 21ATEX0160X-00, order 977822

#### [17] Special conditions for safe use:

1. The power supply to the cabinets is via a plug and socket connection on the electronic control box. The manufacturer delivers appropriate supply cable.



Certification & Inspection

The "Instructions for use" delivered with the product, incorporates the section "Connection to power" – which must be complied.

The mechanism to secure the connection between supply cable plug and cabinet socket must always be used. Connection shall be done in a non-hazardous area. If the cabinets are electrically connected in a zone classified area, the supply cable must be connected in a terminal box complying with national installation requirements e.g. EN 60079-14

- To secure potential equalization of the unit the appropriate cabinet equipotential bonding must be used in accordance with national installation requirements e.g. EN 60079-14.
- 3. The specified ambient temperature for BioCompact II cabinet is:

+10 °C <  $t_{amb}$  < +35 °C (solid door) +10 °C <  $t_{amb}$  < +32 °C (glass door)

- 4. The equipment shall only be used in an area of at least pollution degree 2, as defined in IEC 60664-1 (Indoor laboratories / Office environment).
- 5. Transient protection shall be provided that is set at a level not exceeding 140 % of the peak rated voltage value at the supply terminals to the equipment".
- 6. Due to static hazards cleaning must be done with a damp cloth.

#### [18] Essential Health and Safety Requirements

Concerning EHSR this schedule only verifies compliance with the Directive 2014/34/EU of 26<sup>th</sup> February 2014. The manufacturer's Declaration of Conformity declares compliance with other relevant directives.

#### [19] Drawings and Documents

Scheduled Documents:

Documentation with relevance for the Ex-safety of the product.

Title	Document no.	Date	
Risk analysis	RAC_BC, BCII, BIOBASIC_RR-RF_rev002.xlsm	13 <sup>th</sup> December 2021	
Description of the	Description of BioCompact II 210-310-410	7 <sup>th</sup> December 2021	
product	and variants_002.docx		
Description of PE /	141-2021_Bonding BCII, BIOBASIC.docx	10 <sup>th</sup> December 2021	
bonding connections			
Instructions for use	765042589_20211208_BioCompactII-	10 <sup>h</sup> January 2022	
	Instruction for use_rev001.pdf	- she a	
DoC (Declaration of	BioCompact II 210-310-410 DK	6 <sup>th</sup> December 2021	
Conformity)	06.12.2021.docx		
	BioCompact II 210-310-410 UK		
	06.12.2021.docx		
Measurement of surface	130-2021_Temperature classification of	14 <sup>th</sup> December 2021	
temperatures	BioCompact II		
	RR & RF 210-310-410.docx		



Certification & Inspection

The documentation and scheduled drawings specified for the BioCompact II are stored in the manufacturers "CCD homepage" version dated 06-01-2022. This "CCD homepage" contains information related to several other series of freezers and refrigerators. Only the information related to the type variants listed under item 15 of this certificate are listed as "Scheduled Drawings" covered by this certification.

#### **Additional information**

This certificate has an expiry date on 2024-01-10

Taastrup, 2022-01-10

Danish Technological Institute Certification & Inspection

Jakob Nittegaard ATEX Manager



This certificate must be reproduced in its entirety only and without any changes, including appendices. Project No.: 21.0160 Version: 00